RESULTS OF SELFMONITORING ON GLUCOMETER SYSTEMS ADVANCE AND OPTIUM IN DAILY ROUTINE

Rudolf Chlup^{a,b}, Martin Payne^c, Jana Zapletalová^d, Stanislav Komenda^d, Blanka Doubravová^e, Martina Řezníčková^a, Ludmila Chlupová^f, Pavel Sečkař^g

^a Institute of Physiology, Faculty of Medicine, Palacký University, Olomouc

^b IInd Dept. of Medicine, Faculty of Medicine, Palacký University & Teaching Hospital, Olomouc

^c Hypoguard Limited, Woodbridge, United Kingdom

^d Department of Biometrics, Faculty of Medicine, Palacký University, Olomouc

^e Institute of Neurology and Geriatrics, Moravský Beroun, Czech Republic

^f Department of Special Education, Pedagogical Faculty, Palacký University, Olomouc

^g Department of Health Insurance, Teaching Hospital Olomouc, Czech Republic

Received: April 10, 2005; Accepted (with revisions): May 18, 2005

Key words: *Glucometer Advance/Glucometer Optium/Selfmonitoring/Precision/Software Glucobalance/Software PC-Link*

The aim of this prospective clinical study was to compare the results of B-glucose estimations performed simultaneously on glucometer Advance (with Micro-draw strips) and Optium (G3 strips) by lay healthy volunteers under non-standardized conditions of everyday life, to assess the difficulties dealing with lay – handling of these systems and to demonstrate the possibilities of the software Glucobalance (Hypoguard) and PC-Link (Medisense/Abbott) for the analysis of selfmonitoring. In the course of 5 days, a total of 721 pairs of measurements were carried out on 10 pairs of glucometer Advance and Optium by 10 healthy volunteers aged 16–40 years. The data transfer of all values into computer from glucometer Advance using the Glucobalance software and from glucometer Optium using the PC-Link was carried out to determine the results. The correlation of B-glucose measured on the glucometer Advance and Optium was strong ($r = 0.73$). Glucometer Advance brings values about 0.21 ± 0.06 mmol/l lower than glucometer Optium. The average difference found within each pairs of glucometers Advance – Optium varied. Nevertheless, these differences are acceptable for routine selfmonitoring. The handling of glucometer Advance is not difficult for lay persons. The Glucobalance software simplifies the result evaluation by each tested person. Even though there are some advantages in comparison with the PC-Link, it should be further developed.

INTRODUCTION

Blood glucose selfmonitoring has become a prerequisite in the successful treatment of both type 1 and type 2 diabetes mellitus^{3,4,13,15,22}. In this regard different types of glucometers are marketed to cover increasing demands^{11,24}. The Glucometer Advance (Hypoguard, Great Britain) and the recently developed new generation of capillary strips (Micro-draw) comprise an approved B-glucose self-monitoring system (GSS) which is ready to use in many countries. Glucometer Optium (Medisense/Abbott) and G 3 strips are already known to produce acceptable results from several years experience¹⁴.

The aim of this prospective clinical study was:

1. to compare the results of B-glucose estimations performed simultaneously on glucometer Advance (with Micro-draw strips) and Optium (G3 strips) by lay healthy volunteers under the non-standardized conditions of everyday life;
2. to assess the difficulties dealing with lay – handling of these systems;

3. to demonstrate the possibilities of the software Glucobalance (Hypoguard) and PC-Link (Medisense/Abbott) for the analysis of selfmonitoring.

METHODS

Tested persons

Ten healthy volunteers aged between 16 and 40 years (students, nurses, laboratory assistants) having no personal experience with GSS Advance or Optium entered the study. Informed consent was obtained from all of them. Everyone received one glucometer Advance, one glucometer Optium and two packages of strips per glucometer (Table 1). All volunteers were given a 30 min training in selfmonitoring. None of the glucometers with the exception of Optium No 6 had been used before.

Study design

All measurements were performed in July 2003 in the course of one week, usually at home, at work, in a hospi-

tal ward or on a trip. The weather conditions were good, temperatures mostly in between 15–30 °C.

Each volunteer was encouraged to perform B-glucose selfmonitoring at breakfast and at dinner, according to the following 7-point schedule: before the meal (time 0) and 15, 30, 45, 60, 90 and 120 min following the start of the meal. In this way, everyone was performing 14 estimations per day.

In selfmonitoring, capillary blood was obtained from a finger by means of a lancing device; the first drop was put on a strip G3 (Optium) and, immediately afterwards, the second drop was put on a Micro-draw strip (Advance). The volunteer registered all results read on the displays of both glucometers in the protocol.

At the end of the one-week test period each volunteer was asked to describe the positive and the negative features of individual tested systems.

Statistical analysis

The data were transferred by means of respective software (Glucobalance or PC-Link) from the glucometers to a PC.

The program MS Excel and the statistical package SPSS, v. 10.1, were used for analyzing the results: t-test for matched observations, Correlation Analyses and Analyses of Variance were applied to evaluate the data. The level of significance $p < 0.05$ was taken.

RESULTS

In the course of 5 days, a total of 721 pairs of measurements were carried out on 10 pairs of glucometer Advance and Optium by 10 volunteers. One measurement showed an extreme BG value of 42.8 mmol/l; as this value exceeded the declared range of glucometer (1.1–33.3 mmol/l), the pair concerned was not considered for statistical analysis.

For the results of measurements from individual glucometers see Table 2. Occasionally there were some high (>14.0 mmol/l) or low (<3.3 mmol/l) B-glucose values measured on both glucometers Advance and Optium. There is no explanation for these unusual values.

There was a strong correlation between the B-glucose values measured on all glucometers Advance and Optium ($r = 0.73$, see Fig. 1). However, the correlation coefficients of the individual pairs of glucometers differed ($r = 0.58$ to $r = 0.91$, see Table 3).

For the frequency of B-glucose differences in all 10 pairs of glucometers (Advance – Optium) see the histogram in Fig. 2. The majority of all differences (96.1 %) was within ± 2.25 mmol/l. Extreme differences were registered only rarely (Table 4).

For the results of statistical analysis see Table 5: the B-glucose values on the Advance glucometers were lower about 0.21 ± 0.06 mmol/l ($\bar{x} \pm \text{SE}$), i.e. 2.998 %, than the values measured on the glucometer Optium. This difference did not depend on the absolute B-glucose concentration (Fig. 3).

Table 1. Characteristics of tested glucometer systems Advance and Optium

No	ADVANCE serial number	Strips Batch No	OPTIUM serial number	Strips Batch No
1	SN 040001227	6004	QA0849-4292 P/N 120-203-08	51322
2	SN 040002678	6004	QA2132-1992 P/N 120-503-08	51322
3	SN 040002667	6004	QA2501-1466 120-503-08 C4	51322
4	SN 040002688	6004	QA0849-4532 P/N 120-203-08	51322
5	SN 040002685	6004	QA2501-3595 120-503-08 C4	51322
6	SN 040002689	6004	QA0369-2065 P/N 120-230-08	51322
7	SN 040002704	6004	QA2501-3261 120-503-08 C4	51322
8	SN 040002694	6004	QA2501-1461 120-503-08 C4	51322
9	SN 040002697	6004	QA2202-3555 P/N 120-503-08	51322
10	SN 040002666	6004	QA2202-1771 P/N120-503-08	51322

Table 2. Results of measurements from individual pairs of glucometers Advance (A) and Optium (O)

	PROBAND	ADVANCE	OPTIUM	A-O
1	N	97	97	97
	Minimum	3.3	4.00	-3.09
	Maximum	15.1	13.00	5.16
	Median	8.400	7.5600	.8300
	Mean	8.493	7.5919	.9009
	Std. Deviation	2.561	2.2967	1.5508
2	N	104	104	104
	Minimum	1.4	3.56	-4.79
	Maximum	10.0	10.89	2.56
	Median	5.600	5.9700	-.5900
	Mean	5.565	6.0994	-.5340
	Std. Deviation	1.255	1.0781	1.0807
3	N	56	56	56
	Minimum	3.6	4.61	-1.62
	Maximum	10.0	9.78	2.19
	Median	6.850	5.8900	.7300
	Mean	6.827	6.1836	.6432
	Std. Deviation	1.127	1.2812	.9365
4	N	55	55	55
	Minimum	1.9	3.44	-5.88
	Maximum	12.3	12.50	.69
	Median	5.300	7.2200	-1.6800
	Mean	5.338	7.1251	-1.7869
	Std. Deviation	2.258	2.1265	1.4753
5	N	64	64	64
	Minimum	1.2	3.72	-5.67
	Maximum	21.6	13.39	9.82
	Median	6.050	6.9400	-1.1300
	Mean	6.470	7.4472	-.9769
	Std. Deviation	3.068	1.8851	2.1749

	PROBAND	ADVANCE	OPTIUM	A-O
6	N	100	100	100
	Minimum	3.0	3.17	-3.47
	Maximum	13.9	14.50	1.83
	Median	6.50	6.9150	-.5600
	Mean	6.96	7.5274	-.5684
	Std. Deviation	2.00	2.0497	.8567
7	N	47	47	47
	Minimum	1.1	4.06	-4.70
	Maximum	11.3	10.83	1.13
	Median	5.00	6.1700	-1.0800
	Mean	5.31	6.6472	-1.3366
	Std. Deviation	1.63	1.6876	1.1886
8	N	62	62	62
	Minimum	2.5	4.33	-3.39
	Maximum	11.7	10.83	3.93
	Median	6.80	6.6150	-.1900
	Mean	6.70	6.6485	.0482
	Std. Deviation	1.65	1.4717	1.1254
9	N	78	78	78
	Minimum	3.7	3.28	-1.63
	Maximum	18.8	12.39	8.02
	Median	6.45	6.5600	.1150
	Mean	7.04	6.7637	.2760
	Std. Deviation	2.34	1.9786	1.3998
10	N	57	57	57
	Minimum	3.9	3.67	-1.34
	Maximum	9.0	9.39	2.76
	Median	6.50	5.5600	.6800
	Mean	6.66	5.9556	.7058
	Std. Deviation	1.30	1.5020	.9407
Total	N	720	720	720
	Minimum	1.1	3.17	-5.88
	Maximum	21.6	14.50	9.82
	Median	6.30	6.5600	-.2800
	Mean	6.64	6.8471	-.2053
	Std. Deviation	2.23	1.8838	1.5462

The results of analysis of variance of the differences of each glucometer is shown in Table 6: mean B-glucose differences within individual pairs of glucometers are often significantly different.

The volunteers agreed that the usage of both types of glucometers was simple. The glucometer Advance was appreciated for its small, handy size, the option of going through the memory results forwards and backwards and the very small amount of blood needed for one measurement. On the other hand, the large strip packages didn't fit into the box together with the glucometer and this was seen as a disadvantage.

The data transfer of all values into computer from glucometer Advance using the Glucobalance software and from glucometer Optium using the PC-link was carried out by a professional. The Glucobalance and PC-link enabled as to see the results of an individual volunteer separately. One of the advantages of the Glucobalance is the ability to edit histograms (Fig. 4), pie charts (Fig. 5) and statistics (Fig. 6) for individually set day time periods. The modal day (Fig. 7) and time trends (Fig. 8) are similar to those of the PC Link. None of the programs enables to compare the data interindividually. None of the programs comes in a Czech version.

Table 3. Correlations between the B-glucose values measured on individual pairs of glucometers
Advance (A) – Optium (O)

	O1	O2	O3	O4	O5		O6	O7	O8	O9	O10
A1 Pearson Correlation Sig. (2-tailed) N	.801* .000 97					A6 Pearson Correlation Sig. (2-tailed) N	.911* .000 100				
A2 Pearson Correlation Sig. (2-tailed) N		.580* .000 104				A7 Pearson Correlation Sig. (2-tailed) N		.744* .000 47			
A3 Pearson Correlation Sig. (2-tailed) N			.704* .000 56			A8 Pearson Correlation Sig. (2-tailed) N			.746* .000 62		
A4 Pearson Correlation Sig. (2-tailed) N				.775* .000 55		A9 Pearson Correlation Sig. (2-tailed) N				.802* .000 78	
A5 Pearson Correlation Sig. (2-tailed) N					.712* .000 64	A10 Pearson Correlation Sig. (2-tailed) N					.784* .000 57

* Correlation is significant at the 0.01 level (2-tailed).

Table 4. Frequency table: B-glucose difference Advance – Optium

Difference [mmol/l]	Frequency	Percent	Cumulative Percent	Difference [mmol/l]	Frequency	Percent	Cumulative Percent
Valid -6.25 <= x < -5.75	1	.1	.1	2.25 <= x < 2.75	12	1.7	97.8
-5.75 <= x < -5.25	3	.4	.6	2.75 <= x < 3.25	4	.6	98.3
-5.25 <= x < -4.75	2	.3	.8	3.25 <= x < 3.75	3	.4	98.8
-4.75 <= x < -4.25	5	.7	1.5	3.75 <= x < 5.25	2	.3	99.0
-4.25 <= x < -3.75	2	.3	1.8	4.25 <= x < 4.75	1	.1	99.2
-3.75 <= x < -3.25	5	.7	2.5	4.75 <= x < 5.25	3	.4	99.6
-3.25 <= x < -2.75	11	1.5	4.0	5.25 <= x < 5.75	0	.0	99.6
-2.75 <= x < -2.25	18	2.5	6.5	5.75 <= x < 6.25	1	.1	99.7
-2.25 <= x < -1.75	38	5.3	11.8	6.25 <= x < 6.75	0	.0	99.7
-1.75 <= x < -1.25	51	7.1	18.9	6.75 <= x < 7.25	0	.0	99.7
-1.25 <= x < -0.75	110	15.3	34.2	7.25 <= x < 7.75	0	.0	99.7
-0.75 <= x < -0.25	123	17.1	51.3	7.75 <= x < 8.25	1	.1	99.9
-0.25 <= x < 0.25	115	16.0	67.2	8.25 <= x < 8.75	0	.0	99.9
0.25 <= x < 0.75	72	10.0	77.2	8.75 <= x < 9.25	0	.0	99.9
0.75 <= x < 1.25	63	8.8	86.0	9.25 <= x < 9.75	0	.0	99.9
1.25 <= x < 1.75	40	5.6	91.5	9.75 <= x < 10.25	1	.1	100.0
1.75 <= x < 2.25	33	4.6	96.1	Total	720	100.0	

Table 5. Statistical evaluation of the B-glucose difference Advance – Optium (all results)**One-Sample Statistics**

	N	Mean	Std. Deviation	Std. Error Mean
A_O Difference	720	– .205	1.5461	.0576

One-Sample Test

	Test Value = 0					
	t	df	Sig. (2-tailed)	Mean Difference	95% Confidence Interval of the Difference	
					Lower	Upper
A_O Difference	–3.562	719	.0004	– .205	– .318	– .092

Table 6. Statistical evaluation of the B-glucose difference Advance – Optium (ANOVA – multiple comparisons of individual patients i.e. pairs of glucometers)

(1) PROBAND	(J) PROBAND	Mean Difference (I–J)	Std. Error	Sig.	95% Confidence Interval	
					Lower Bound	Upper Bound
1	2	1.435*	.186	.000	1.070	1.800
	3	.258	.221	.244	–.176	.691
	4	2.688*	.222	.000	2.252	3.124
	5	1.878*	.212	.000	1.462	2.294
	6	1.469*	.188	.000	1.101	1.838
	7	2.238*	.234	.000	1.778	2.697
	8	.853*	.214	.000	.432	1.273
	9	.625*	.200	.002	.232	1.018
	10	.195	.220	.375	–.236	.626
2	3	–1.177*	.218	.000	–1.606	–.749
	4	1.253*	.219	.000	.822	1.684
	5	.443*	.209	.035	.032	.853
	6	.034	.184	.852	–.328	.396
	7	.803*	.231	.001	.348	1.257
	8	–.582*	.211	.006	–.997	–.168
	9	–.810*	.197	.000	–1.197	–.423
	10	–1.240*	.217	.000	–1.666	–.814

* The mean difference is significant at the .05 level.

Table 6 cont.

(1) PROBAND	(J) PROBAND	Mean Difference (I-J)	Std. Error	Sig.	95% Confidence Interval	
					Lower Bound	Upper Bound
3	4	2.430*	.250	.000	1.940	2.921
	5	1.620*	.241	.000	1.147	2.093
	6	1.212*	.220	.000	.780	1.643
	7	1.980*	.260	.000	1.469	2.491
	8	.595*	.243	.014	.119	1.071
	9	.367	.231	.112	-.085	.820
	10	-.063	.248	.801	-.549	.424
4	5	-.810*	.242	.001	-1.285	-.335
	6	-1.219*	.221	.000	-1.652	-.785
	7	-.450	.261	.085	-.964	.063
	8	-1.835*	.244	.000	-2.314	-1.356
	9	-2.063*	.232	.000	-2.518	-1.608
	10	-2.493*	.249	.000	-2.981	-2.004
5	6	-.408	.211	.053	-.822	.005
	7	.360	.253	.155	-.137	.856
	8	-1.025*	.235	.000	-1.486	-.565
	9	-1.253*	.222	.000	-1.689	-.817
	10	-1.683*	.240	.000	-2.153	-1.212
6	7	.768*	.233	.001	.311	1.225
	8	-.617*	.213	.004	-1.034	-.199
	9	-.844*	.199	.000	-1.235	-.454
	10	-1.274*	.218	.000	-1.703	-.845
7	8	-1.385*	.255	.000	-1.885	-.885
	9	-1.613*	.243	.000	-2.090	-1.135
	10	-2.042*	.259	.000	-2.552	-1.533
8	9	-.228	.224	.309	-.668	.212
	10	-.658*	.242	.007	-1.132	-.183
9	10	-.430	.229	.061	-.880	.021

* The mean difference is significant at the .05 level.

Fig. 1. Correlation between the B-glucose values measured on all glucometers Advance and Optium

Fig. 3. Correlation between the B-glucose difference Advance - Optium and the B-glucose value measured on glucometer Optium

Fig. 2. Histogram: frequency (number of observations) of individual B-glucose differences Advance - Optium; total number of comparisons $n = 720$

Fig. 4. Histograms of B-glucose values from the defined day time periods in one tested person (GSS Advance, software Glucobalance)

Fig. 5. Pie charts showing the frequency of B-glucose values in defined ranges in the course of defined day time periods (GSS Advance, software Glucobalance)

Display	1	weeks ending	20. 7. 2003	Data available from 1.1.2000 to 9.10.2003					
	Pre-Dawn 0:00-6:59	Breakfast 7:00-8:59	Snack AM 9:00-10:59	Lunch 11:00-13:59	Snack PM 14:00-16:59	Dinner 17:00-18:59	Snack Eve 19:00-21:59	Night 22:00-23:59	Aggregate
# Readings	0	8	20	1	0	23	18	0	70
Maximum		17,37	21,59	5,16		8,10	9,49		21,59
75th Percentile		8,30	8,48	5,16		6,71	6,83		6,96
Median		5,27	6,05	5,16		6,10	5,69		5,99
25th Percentile		4,33	4,95	5,16		4,63	4,58		4,56
Minimum		1,05	3,05	5,16		1,05	1,05		1,05
Mean		7,28	7,48	5,16		5,57	5,54		6,30
Std. Dev.		5,59	4,49	0,00		2,01	2,16		3,47
Events									
Hyper (>=10,0)		2	4	0		0	0		6
Hypo (<=3,0)		1	0	0		3	2		6
Above Target		25%	2 30%	6 0%	0	9%	2 11%	2	17%
On Target		63%	5 55%	11 100%	1	78%	18 67%	12	67%
Below Target		13%	1 15%	3 0%	0	13%	3 22%	4	16%

Fig. 6. Statistics (GSS Advance, software Glucobalance)

Fig. 7. Modal day: all B-glucose values measured in the course of several days drawn into the time coordinates of one day (GSS Advance, software Glucobalance)

Fig. 8. Trend chart: all B-glucose values drawn in chronological order (GSS Advance, software Glucobalance)

Fig. 9. Clarke Error Grid Analysis of Capillary finger-prick data from persons with type 1 diabetes; investigations were performed in Ipswich Hospital Diabetic clinics, Great Britain (2003–2004) (ref.²⁰)

DISCUSSION

The results of B-glucose selfmonitoring depend on the quality of the glucometer, on the quality of strips and on various preanalytical and analytical factors such as temperature, humidity¹⁷, amount of blood¹⁹, B-glucose concentration and other physical-chemical parameters (including the presence of drugs etc.) influencing the activity of the glucosidase and the size of the developing electrical flow^{6, 7, 8, 16, 18}. The registered B-glucose values collected to check the function of the glucometer system Advance was influenced by all above mentioned factors. In addition to it, the person performing the selfmonitoring is one of the inevitable parts of the whole procedure.

The laboratory regulations dealing with the accuracy of the glucometers demand that the value measured is not different from the values estimated using the approved analyzer more than by 15 % (ref.^{2, 5, 9, 10}). The accuracy of both glucometers Advance and Optium has been evaluated by several researches^{1, 12, 14, 21, 23}.

In this user-oriented study it was not possible to carry out measurements on an approved analyzer and therefore it is not possible to produce regarding evidence which kind of glucometer is more accurate. On the other hand it is possible to conclude that the average difference of B-glucose measured on glucometers Advance differs from the glucometer Optium. The magnitude of difference Advance – Optium does not seem to be dependent on the absolute B-glucose value. However, this assessment is limited, as the estimations were performed only in the B/glucose concentration range from 3.17 to 14.50 mmol/l. The well known glucometer Optium was considered to be “empirically competent” to be used as a reference glucometer; nevertheless, it should be taken into consideration that the differences could be due to inaccuracy and imprecision of both glucometers, i.e., Optium and Advance¹⁴. In addition, the importance of various differences of separate pairs of glucometers is limited, because each pair was tested by a different person and under different circumstances. The outcome for the praxis is that each person should always perform selfmonitoring with the same type of the personal glucometer. Results of individual pairs of glucometers are shown (Table 1 and Table 2) to demonstrate the variability of estimations as it may be expected by the physician in daily routine.

The software Glucobalance was not used in the Czech Republic before the year 2003 and no references describing some experience with it could be found. Therefore, Fig. 4–8 was included into this paper to mention and assess its advantages and disadvantages.

In a recent Advance Micro-draw study²⁰, the quality of the GSS was evaluated using persons with diabetes attending Ipswich Hospital Diabetic clinics from December 2003 till January 2004. Capillary blood was obtained directly from finger-pricks. Results were compared to a YSI reference analyser (Fig. 9). All data is presented without any recognised outlier removals. Very low (< 2.8 mmol/l) glucose results were obtained using capillary blood that was incubated to lower glucose content prior to meter

measurement. Data was evenly stratified across the blood glucose measurement range. There was no screening of patients attending the diabetic outpatients clinic. There was no limitation made on drug therapy/interferences that may have been present in the blood and no limitation on the hematocrit of any patient entering the study. (A total of 8 meters were involved in these clinical measurements).

In conclusions in may be said:

1. The results of B-glucose selfmonitoring by means of both tested glucometer systems Advance and Optium under non-standardized conditions of daily life appears to be reliable. The correlation of B-glucose measured on the glucometer Advance and Optium was strong ($r = 0.73$). Glucometer Advance produce values about 0.21 ± 0.06 mmol/l lower than glucometer Optium. The average difference found within each pairs of glucometers Advance – Optium varied, nevertheless, these differences are acceptable for the routine self-monitoring.
2. The handling of glucometer Advace is not difficult for lay persons.
3. The Glucobalance software simplifies the result evaluation by each tested person. Even though there are some advantages in comparison with the PC-Link, it should be further developed.

ACKNOWLEDGEMENTS

This study was supported by research projects IGA Nr 7825-3, Ministry of Health, MSM 151100005 and Ministry of Education Youth and Sports, Czech Republic.

REFERENCES

1. Alamia V Jr, Jaekle R, Royek A, Meyer BA. (1998) Accuracy of portable glucose meters for rapid determination of amniotic fluid glucose levels. *J Maternal – Fetal Med* 7, 183–189.
2. American Diabetes Association. (2002) Tests of glycemia in diabetes. *Diabetes Care* 25, 597–599.
3. American Diabetes Association. (1994) Self-monitoring of blood glucose (Consensus Statement). *Diabetes Care* 17, 81–86.
4. Dillon AE, Menard MK, Rust P, Newman RB, VanDorsten JP. (1997) Glucometer analysis of one-hour glucose challenge samples. *Am J Obstet Gynecol* 177, 1120–1123.
5. Dohnal L, Frýdecký B, Zima T, Kalla K, Škrha J. (2000) Doporučení č. 2 České společnosti klinické biochemie ČLS JEP, Referenční laboratoře pro klinickou biochemii MZ ČR a České diabetologické společnosti: Zabezpečování jakosti měření glukometry při použití v režimu point-of-care testing. *Klin Bioch Met* 9/30, 78–79.
6. Dybkaer R. (1999) From total allowable error via metrological traceability to uncertainty of measurement of unbiased results. *Accred Qual Assur* 6, 401–405.
7. Fazel A, Koutoubi Z, Sorg TB, Mehrotra B. (1996) Influence of sample temperature on reflectance photometry and electrochemical glucometer measurements. *Diabetes Care* 10, 771–774.

8. Fink KS, Christensen DB, Ellsworth A. (2002) Effect of high altitude on blood glucose meter performance. *Diab Technol Therap* 4, 627–635.
9. Fogh-Andersen N, D'Oratio P, Kunze K, Kuelpmann W, Mager G, Latson L. Recommendation on reporting results for blood glucose. <http://www.ifcc.org/ejifcc/voll2no4/voll2no4a4.htm>
10. Frýdecký B. (2001) Laboratorní diagnostika diabetu a požadavky na měření koncentrace glukózy. *Klin Bioch Met* 9/30, 173–177.
11. Goldstein DE, Little RR, Lorenz RA, Malone JI, Nathan D, Peterson CM. (1995) Test of glycemia in diabetes (Technical Review). *Diabetes Care* 18, 896–909.
12. Haag BL, Leed I. (1999) Suceptibility of two new glucose test strips to testing errors. *Diabetes* 48 (Suppl. 1), A415.
13. Chee F, Fernando T, van Heerden PV. (2003) Closed-loop glucose control in critically ill patients using continuous glucose monitoring systém (CGMS) in real time. *IEEE Transaction on Information Technology in Biomedicine* 7, 43–53.
14. Chlup R., Bartek J, Malá E, Doubravová B, Pukowietz L, Zatloukal P, Chlupová L, Zapletalová J. (2004): Uživatelská studie o správnosti a přesnosti měření glukometrů Advance, Card a Optium (User-oriented study on accuracy and precision of glucometer systems Advance, Card and Optium). *Klin Bioch Metabol* 12, 171–178.
15. Karter AJ, Ackerson LM, Darbinian JA, D'Agostino RB Jr, Ferrara A, Liu J, Selby JV. (2001) Self-monitoring of blood glucose levels and glycemic control: the Northern California Kaiser Permanent Diabetes Registry. *Am J Med* 111, 1–9.
16. Kellner A, Waldenstroem J. (1999) Estimation of total uncertainty of glucose measurements. *Clin Chem Lab Med* 37, Spec. Suppl. S14.
17. King JM, Eigenmann CA, Colagiuri S. (1995) Effect of ambient temperature and humidity on performance of blood glucose meters. *Diab Med* 12, 337–340.
18. Kurahashi K, Maruta H, Usuda Y, Ohtsuka M. (1997) Influence of blood sample oxygen tension on blood glucose concentration measured using an enzyme-electrode method (comment). *Critic Care Med* 25, 231–235.
19. Lewandrowaki KB, Dan L. (2002) Effects of small sample volume and interfering substances on two glucose meters (Vliv malého objemu krve v testovaném vzorku a obsahu příměsí na dva měřicí přístroje pro diabetiky). *Diabetes* 48 (Suppl.1), A387.
20. Payne M. (2004) Hypoguard Advance Micro-draw Performance. *Klin Bioch Metabol* 12, 202.
21. Puentmann I, Wosniok W, Haeckel R. (2003) Comparison of several point-of-care testing (POCT) glucometers with an established laboratory procedure for the diagnosis of type 2 diabetes using the discordance rate. A new statistical approach. *Clin Chem Lab Med* 41, 809–820.
22. Rumley AG (1997) Improving the quality of near-patient blood glucose measurement (comment). *Annals Clin Bioch* 34 (Pt 3), 281–286.
23. Stahl M, Jorgensen LGM, Hyltoft Petersen P, Branslund I et al (2001) Optimization of preanalytical conditions and analysis of plasma glucose. 1. Impact of the new WHO and ADA recommendations on diagnosis of diabetes mellitus. *Scand J Clin Lab Invest* 61, 169–180.
24. Thai AC, Ng WY, Lui KF, Yeo PP (1989) Three new glucose reflectance meters: Diascan, Glucometer II, and Reflolux II. *Diabetes Res Clin Pract* 7, 75–81.